

VLBA ACQUISITION MEMO #336

MASSACHUSETTS INSTITUTE OF TECHNOLOGY
HAYSTACK OBSERVATORY
WESTFORD, MASSACHUSETTS 01886

21 August 1992

Telephone: 508-692-4764

Fax: 617-981-0590

To: VLBA Data Acquisition Group
From: Dan Smythe
Subject: 16 Passes with a 2 μm Guard Band

VLBA Acquisition Memo #119 reported good performance with a 5 μm guard band. I have discovered that peaking is possible, and error rates are low, even with a negative guard band. To better understand what happens when tracks are put close together, I measured the total power at 160 ips, for an isolated track and for a track with adjacent tracks recorded at $\pm 40 \mu\text{m}$. The results are shown in Fig. 1, where the solid curves are sums of squared triangles, since the recorded waveforms are uncorrelated. Note that the power drops 3 dB between tracks and provides a good signal for peaking, even at offsets as small as 3 μm .

Error rates for two tracks on these two recordings are shown in Fig. 2. For the narrow guard band, errors start to rise at an offset of 4 to 6 μm , and time is not decoded at offsets greater than $\pm 12 \mu\text{m}$; whereas for an isolated track, errors stay low for offsets as high as 12 μm . The conclusion is that guard bands as low as 2 μm , for passes in the same direction, are possible, provided the peaking algorithm does not stray more than 5 μm from the center of the track.

For passes in opposite directions, the minimum guard band is determined by anisotropies in the tape, as described in VLBA Acquisition Memo #129, and by the accuracy in calibrating the forward-reverse offset. For Mark IIIA Mode A, the reverse guard band is 35.5 μm , and all over-writing noticed at the correlator has been the result of calibration problems, rather than tape anisotropies.

A track spacing of 45 μm would provide 14 passes with a same-direction guard band of 7 μm and an opposite-direction guard band of 41 μm . A track spacing of 40 μm would provide 16 passes with a same-direction guard band of 2 μm and an opposite-direction guard band of 31 μm , which is probably acceptable, provided the recorders are kept well-calibrated and well-maintained.




Figure 1. Reproduce Power as a function of head offset.

