

NATIONAL RADIO ASTRONOMY OBSERVATORY
CHARLOTTESVILLE, VIRGINIA

ELECTRONICS DIVISION INTERNAL REPORT No. 255

A SYNTHESIZED, 90-120 GHz SIGNAL SOURCE

J. W. ARCHER AND M. J. CRAWFORD, JR.

DECEMBER 1984

NUMBER OF COPIES: 150

A SYNTHESIZED 90-120 GHZ SIGNAL SOURCE

J. W. Archer and M. J. Crawford, Jr.

Table of Contents

I.	Introduction	1
II.	The 26.5-40 GHz Frequency Synthesizer	2
	A. General Description	2
	B. Design of the UHF Synthesizers	5
	C. The Coarse Tuning and Bias Modulation Amplifiers	11
	D. The Main 26.5-40 GHz Phase Lock Loop	14
III.	The 90-120 GHz Frequency Tripler	21
IV.	The Complete Synthesized LO Source	27
V.	Summary	27
VI.	Acknowledgements	28

Figures

Figure 1	26.5-40 GHz Synthesizer Block Diagram	3
Figure 2	0.88-1.32 GHz Synthesizer Schematic	7
Figure 3	ROM and 9's Complement Encoder Schematic	9
Figure 4(a)	Spectrum of 1.76-2.64 GHz Synthesizer Output	10
	(b) Spectrum of 1.76-2.64 GHz Synthesizer Output	10
	(c) 400-500 MHz Reference Signal Spectrum	12
	(d) I. F. Spectrum in Main Loop	12
Figure 5	Tuning Coil Driver Schematic	13
Figure 6	Harmonic Mixer Triplexer Schematic	15
Figure 7	Sweep Generator Schematic	18
Figure 8	FM Coil Driver and Loop Amplifier Schematic	20
Figure 9(a)	Tripler Block: Cross-Sectional View	22
Figure 9(b)	Tripler Block: Cross-Sectional View	23
Figure 10	Output Power vs Output Frequency for Tripler Alone and for Complete Signal Source	25
Figure 11	Tripler Conversion Efficiency vs Input Power	26
References	29

A SYNTHESIZED 90-120 GHZ SIGNAL SOURCE

J. W. Archer and M. J. Crawford, Jr.

I. Introduction

The advantages of all-solid-state millimeter wavelength receivers have been clear to system designers for many years. Substantial savings in cost and weight, and significantly enhanced system reliability, can be achieved with a completely solid-state design. Until recently, achieving a widely tuneable solid-state local oscillator system has proven to be a significant obstacle to the development of broadband heterodyne receiver systems of this type. Developments in millimeter wave harmonic generator design [1],[2], Gunn oscillator technology [1],[3] and the reduction in the L0 power requirements for heterodyne receivers through the use of improved, cooled Schottky diode [4],[5] or superconducting mixer [6],[7] elements have now made it feasible to realise practical receivers with solid state L0's. In many applications, frequency and/or phase stability of the L0 source is of prime importance. The construction of a highly stable, widely tuneable synthesizer [8], such as described here, is therefore a significant development.

The following sections of this paper will describe the design and construction of the 26.5-40 GHz frequency synthesizer. Performance data will be given, which illustrate the frequency stability, output power and spectral purity of the phase locked source. The design of the frequency tripler is then described and performance curves are given. Finally, data is presented showing the performance of the complete synthesized L0 source.

II. The 26.5-40 GHz Frequency Synthesizer

A. General Description:

The principal millimeter wavelength signal source in the LO system is a YIG tuned GaAs Gunn oscillator manufactured by Watkins Johnson, Inc. (Model WJ-5610-302F). This oscillator may be electronically tuned between 26.5 and 40 GHz by varying the magnetic field applied to a spherical YIG resonator to which the Gunn diode is coupled. The magnetic field is produced by current flowing in the coils of an electromagnet, whose pole pieces enclose the YIG element. As the tuning coil current is varied linearly from 660 mA to 1015 mA, the oscillator frequency varies in a smooth, almost linear response from 26.5 to 40 GHz. The peak deviation from linearity is about -60 MHz near the upper and lower tuning limits. The tuning coil resistance is about 10 ohms with an inductance of 160 mH. Another electromagnetic circuit is provided to enable fine tuning of the oscillator frequency. This coil gives a sensitivity of 125 kHz/mA with a maximum current of ± 800 mA, resulting in a peak attainable deviation of ± 100 MHz. In order to achieve the flattest output power response as the output frequency is varied, the supply voltage to the Gunn diode must be changed linearly and in synchronism with the tuning coil current - at 26.5 GHz the optimum Gunn bias is 5.2 volts and at 40 GHz the optimum bias is 3.4 volts. When the bias voltage is controlled in this manner, the output power over the full tuning range is +13 dBm ± 1 dB.

Figure 1, a block diagram of the 26.5 to 40 GHz synthesizer system, shows how the YIG tuned oscillator is incorporated into the phase lock system. The coarse tuning coil current and Gunn diode bias are derived by digital to analog conversion of the BCD-encoded frequency select input, so that the initial oscillator frequency is set as near as possible to the desired output frequency. An automatic

fine frequency sweep signal can be injected at the input of the coarse tuning coil bias amplifier in order to tune the oscillator close enough to the desired frequency for lock acquisition. The microwave output from the YIG oscillator is fed through a broadband waveguide isolator and then sampled by a 20 dB cross-guide coupler. The sampled signal is fed to a harmonic mixer (Hewlett-Packard Model 11517A) where it is mixed with the 15th harmonic of a 1.76-2.64 GHz reference signal. This signal is generated by extracting the second harmonic of the output of one of two UHF frequency synthesizers in the system, both of which are locked to a common, highly stable, 10 MHz reference signal. The frequency of the 2 GHz reference is controlled by the first three digits of the frequency select input data. Decimal 4 is subtracted from this number to correct for the 400 MHz offset of the IF reference synthesizer and its 9's complement code generated, in order to derive the required preset data to determine the division modulus for the programmable divider in the UHF synthesizer. A numerical example will help in understanding the relationships:

Let the selected frequency be 31.6, 543 GHz (no special significance). The division modulus will be $316 - 4 = 312$. The output frequency from the UHF synthesizer will be $(10 \text{ MHz}/12) \times 312 \times 4 = 1040 \text{ MHz}$. Its second harmonic component, which is fed to the harmonic mixer, will be 2080 MHz. When the YIG oscillator is tuned to the desired frequency, the IF output frequency from the harmonic mixer will be $31,654.3 - 15 \times 2080 = 454.3 \text{ MHz}$.

The IF output from the harmonic mixer is amplified and fed to a phase detector where it is correlated with a signal derived from the second UHF synthesizer. The frequency of the second synthesizer is controlled by the remaining three digits of the frequency select command and incorporates a fixed frequency offset of 400 MHz. Thus, in the example above the output frequency would be

$400 + 54.3 = 454.3$ MHz, which is exactly the required frequency if the YIG tuned oscillator is to lock at 31.6543 GHz. The output of the phase detector is filtered, amplified and fed back to control the phase of the YIG oscillator via the bias applied to the fine tuning (FM) coil.

The use of separate synthesizers for coarse and fine control of the reference frequencies is essential if acceptable phase noise is to be achieved in the spectrum of the locked millimeter wave oscillator. The phase noise of the output signal, within a loop bandwidth separation from the carrier, will be determined primarily by the phase noise characteristics of the harmonic mixer reference signal. If only the 1.76-2.64 GHz reference is made variable and the other reference is fixed at, say, 400 MHz, it would be necessary to operate the phase detector at 833 Hz in the remaining synthesizer to achieve a 100 KHz increment in the millimeter wave output signal. Since the carrier-to-noise ratio for phase noise degrades as the square of the frequency ratio, the FM noise characteristics of the primary reference oscillator would be degraded by about 70 dB in the final output spectrum - clearly an unacceptable situation for practical oscillators. By controlling the frequency of both reference signals and carrying out fine frequency adjustment using an IF frequency offset, the phase noise degradation is reduced to about 40 dB, which is acceptable when a high quality 10 MHz reference oscillator is employed.

B. Design of the UHF Synthesizers:

The 1.76-2.64 GHz and 400-500 MHz synthesizers are very similar in design, using essentially the same digital circuitry for prescaling, variable modulus division and phase/frequency detection. The significant differences are the different reference frequencies (833 kHz vs 25 kHz), the number of variable modulus counter stages (3 vs 4) and the manner in which the frequency select

digits are derived. Because of the similarities, only the 1.76-2.64 GHz synthesizer circuit will be described in detail.

In this case, because of the technological constraints on prescaler performance, the phase lock loop operates at half the desired output frequency, i.e., 0.88-1.32 GHz. The oscillator employed is a varactor tuned transistor oscillator (EMF VTO-4B) which provides more than +13 dBm over the fundamental frequency range and -25 dBm \pm 2 dB at the second harmonic. The output from the oscillator is split by a 3 dB power divider. One output from the divider is further attenuated by 10 dB and fed to a divide by 4, ECL prescaler. The other output is passed through a bandpass filter, which passes the desired second harmonic component but rejects the fundamental and other higher order harmonics from the oscillator. The filter output is fed to a variable gain, broadband power amplifier chain with a total maximum gain of 40 dB, boosting the signal power to a nominal maximum of +15 dBm. A second identical bandpass filter follows, before the reference signal is coupled to the harmonic mixer. The second filter prevents noise from the power amplifiers reaching the high gain, low noise IF chain of the millimeter wave phase lock loop.

The fixed modulus, divide by 4, ECL prescaler (Plessey SP8611B) at the input of the digital divider chain is followed, as shown in Figure 2, by a dual modulus ($\div 10/\div 11$) ECL counter (11C90). This counter is interconnected with a series of synchronously clocked, presetable, high speed TTL BCD decade counters (74F160) to implement a variable modulus counter using pulse swallowing techniques [9]. The use of the Schmitt trigger (74S14) to generate the synchronous clock signal for the 74F160 counters helps eliminate false clocking problems encountered at the higher input frequencies. Printed circuit, 50 ohm microstriplines are used to couple the high frequency signals to the SP8611B and 11C90 counters.




Fig. 2. A detailed circuit diagram of the 0.88-1.32 GHz digital frequency synthesizer. The 1.76-2.64 GHz reference signal fed to the millimeter wave harmonic mixer is obtained by filtering and amplifying the second harmonic of the VCO output signal.

The coupling and bypass capacitors are chip capacitors. The bypass capacitors are grounded to the ground plane of the board using low inductance, plated-through holes.

The division modulus preset digits to the programmable counters are derived using 9's complement encoders; one channel is illustrated in Figure 3. The subtract decimal 4 function is performed prior to 9's complement encoding using a look-up table stored in a set of read only memories. The complete circuit and table are given in Figure 3. The output signal from the programmable divider is fed, via a level translating resistor network, to an ECL phase frequency detector (12040). The 833 kHz reference input for the detector is derived from the 10 MHz standard using a divide by twelve, 74LS92 counter. The output pulses from the phase detector are coupled to an active second order loop filter/integrator with adjustable damping controls. An AM-464 (Analog Devices) operational amplifier is used, which is capable of handling the high voltages required to directly bias the tuning varactor in the VCO for complete coverage of the oscillator tuning range. A common problem with a loop of this type is the appearance of spurious components in the spectrum of the locked oscillator spaced at harmonics of 833 kHz from the desired output frequency. These arise because of insufficient filtering of the sharp output pulses from the phase/frequency detector at the varactor bias output. In the present design, with a loop natural frequency of about 20 kHz [10], RC pulse prefiltering (0.01 μ F capacitors in the input resistor network) and a series resonant LC trap at the op-amp output are used to reduce the amplitude of the spurious signals to acceptable levels. Photographs of the typical output spectrum (at 2 GHz) from the 1.76-2.64 GHz synthesizer, shown in Figures 4(a) and 4(b), demonstrate that the unwanted components are at least 60 dB below the carrier. A simple lock indicator circuit is implemented






Fig. 3. A diagram showing the design of the \overline{BCD} subtract four, ROM circuit and the \overline{BCD} to "9's complement of BCD" converter. The inset table gives the logic used in implementing the look-up table stored in ROM.



a) The 1.76-2.64 GHz reference signal to the harmonic mixer - set at 2.0 GHz. Analyzer B.W. = 3 KHz, horizontal scale = 20 KHz/div, vertical scale = 10 dB/div.



b) As for a) but with horizontal scale = 500 KHz/div.


Fig. 4. Photographs of typical signal spectra at important monitor points in the synthesizer system.

using a differential amplifier to detect the presence of a frequency correction signal from the 12040, followed by a dual threshold comparator which gives a 5 volt output signal when the loop is unlocked.


Figure 4(c) shows the typical output spectrum from the 400-500 MHz synthesizer. The VCO used in this phase locked loop is an Avantek 8040 varactor tuned device. A bandpass filter is required between oscillator and prescaler to eliminate false triggering of the divider due to the second harmonic component at the output of the VCO. In this loop, four 74F160 counters are used to implement the variable modulus counter, but the most significant digit is fixed at a count of 4. The code for the 3 remaining digits is a 9's complement encoded version of the last three digits of the frequency select command input.

C. The Coarse Tuning and Bias Modulation Amplifiers:

Figure 5 gives a schematic diagram of the coarse tuning coil current driver and Gunn diode bias modulator. A pair of three digit BCD, digital-to-analog converters are used to generate an analog voltage from the BCD frequency select input. The resultant tuning voltage at the output of the summing amplifier A is equal to the selected frequency in GHz divided by 10. This voltage drives a linear voltage to current converter (op-amp B and ECG261 transistor) whose gain and offset can be adjusted so that the YIG tuned oscillator frequency is as close as possible to the desired locking frequency. Coil current is sensed by a precision 1 ohm resistor in series with the windings. The output voltage from amplifier A is also used to control the bias voltage applied to the Gunn diode in order to minimize the variations in YIG oscillator output power with operating frequency. The voltage from amplifier A is fed to non-inverting amplifier C, which has a voltage gain of 3.3 and includes an adjustable offset control. The output from this amplifier is fed via a 14.3K resistor to the control input



- c) The 400-500 MHz reference signal to the phase detector quadrature hybrid - set at 450 MHz. Analyzer B.W. = 3 KHz, horizontal scale = 20 KHz/div., vertical scale = 10 dB/div.



- d) The spectrum of the IF signal at the input to the phase detector - YIG oscillator locked at 30.450 GHz. Analyzer B.W. = 30 KHz, horizontal scale = 100 KHz/div., vertical scale = 10 dB/div.

Fig. 4. Photographs of typical signal spectra at important monitor points in the synthesizer system.

BCD
CODE




Fig. 5. A schematic diagram of the coarse tuning coil driver circuit showing the D/A converter, driver amplifier and Gunn diode bias modulator.

of a current controllable, positive 1 amp voltage regulator (LAS15U). The sensitivity of the inverting control input is set by a variable 10K feedback potentiometer connected between the regulator output and ground. By suitably adjusting the gain and offset controls, the Gunn bias voltage can be made to vary linearly from 5.2 to 3.6 volts as the tuning voltage varies from 2.65 to 4.0 volts. The 1N1186 diodes connected between the Gunn diode and the regulator provide a constant voltage drop of 1.2 volts, allowing the LAS15U to operate at output voltages between 6.8 and 4.8 volts, which are above its minimum output voltage limit of 4.0 volts. The zener diode and fuse protect the Gunn diode against overvoltage conditions that might occur in the event of failure of the regulator or other circuit element.

Op-amp D, connected as a comparator, is used to prevent an incorrect input code from causing the application of an excessive bias voltage to the Gunn oscillator. Should a BCD input number be selected which is less than 26.5 GHz, the comparator output swings positive, forcing the tuning voltage to remain constant at 2.65 volts. If the selected frequency is above 40 GHz, then the Gunn voltage will be below 3.5 volts and the tuning coil current is limited to a safe value of less than 1.5 amps by the internal resistance of the windings.

D. The Main 26.5-40 GHz Phase Lock Loop:

As described previously, a sample of the YIG tuned oscillator output is fed to a Hewlett Packard harmonic mixer where it is mixed with the 15th harmonic of a reference signal in the range 1.76-2.64 GHz. The reference signal, mixer bias and IF output signal are coupled to and from the single coaxial port of the mixer using a triplexing filter shown schematically in Figure 6. The triplexer is comprised of a pair of three-element, lowpass filters which isolate the high power, 2 GHz reference signal from the 450 MHz IF amplifiers, and both of these




Fig. 6. A diagram showing the principal features of the harmonic mixer signal triplexer. The diagram also shows the automatic level control system which adjusts the power level of the 1.76-2.64 GHz signal to maintain a constant bias current in the harmonic mixer diode.

signals in turn from the mixer DC bias input. DC blocking capacitors are provided at the RF and IF ports. The triplexing filter was fabricated using microstripline techniques on aluminum-backed, Teflon fiberglass dielectric material.

The harmonic mixer, which has widely varying sensitivity over the range 1.76-2.64 GHz, is operated in a zero voltage bias mode at a constant forward current of 4.8 mA. The bias current is held constant by adjusting the power of the 2 GHz input using a servo loop (shown in Figure 6), which controls the gain of the reference signal power amplifier. Operation at a constant bias current ensures a more uniform sensitivity for the harmonic mixer between 26.5 and 40 GHz and reduces variations in the IF power at the phase detector.

The 400-500 MHz IF output from the triplexer is passed to a lowpass filter (500 MHz cutoff), which further attenuates the 2 GHz reference signal, before being fed to a low noise, high gain amplifier (3 dB NF, 45 dB gain). The first amplifier is followed by a 400-500 MHz bandpass filter, a monitor directional coupler, a limiting amplifier and a 10 dB attenuator, resulting in a nominal input power to the phase detector power divider of +3 dBm. The reference signals for the two double balanced mixer phase detectors are derived from the amplified 400-500 MHz synthesizer output using a quadrature hybrid. Nominal input power to the hybrid is +3 dBm.

In order for the loop to be able to acquire lock, the difference frequency between the IF signal from the harmonic mixer and the phase detector reference signal must be less than the bandwidth of the loop filter (about 500 KHz). This will not be the case when the YIG oscillator frequency is preset by the coarse tuning coil because of the deviation of the tuning current vs frequency response from linearity. Nevertheless, the difference frequency should be less than 60 MHz for the oscillator used here. Provision is made in the design of

the coarse tuning coil driver amplifier for the injection of a small sweep offset voltage, which can be used to bring the YIG oscillator frequency within the lock range of the loop.

The sweep circuit is illustrated in Figure 7. The sweep function is implemented digitally, a technique which has significant advantages over an analog method. A 256 Hz clock pulse generated by a 555 timer is fed via a control gate to a cascade of two 74LS193 binary up/down counters. The counter outputs are coupled to an eight-bit D/A converter, which generates a repetitive voltage ramp as the counters are clocked. The voltage ramp is fed via an appropriate scaling resistor to the input of the coarse tuning coil driver amplifier, resulting in a modulation of the YIG oscillator frequency by about ± 100 MHz in synchronism with the sweep voltage. When the YIG oscillator frequency comes within the capture range of the loop (about 1 MHz) and lock is acquired, a voltage comparator connected to the output of the quadrature phase detector senses the locked condition and generates a +5 volt signal which disables the clock to the sweep counters. The coarse tuning bias is thus latched at the condition at which lock was acquired.

A problem that has been encountered with the YIG oscillator used here is significant frequency vs temperature dependence. In operation, after the loop is locked, ambient temperature changes can cause the oscillator to drift out of the range of the loop correction signal. A simple addition to the sweep circuit has enabled this problem to be overcome. As shown in Figure 7, the current in the YIG oscillator FM coil is converted to a suitable proportional voltage and sensed by a dual threshold comparator with hysteresis, which is triggered should the correction voltage exceed ± 2 volts (available range is ± 4 volts at a sensitivity of 0.1V/A). The comparator outputs control gates which feed the clock signals to the up or down clock inputs of the binary counter




Fig. 7. The ramp generator used to scan the YIG oscillator frequency until lock is acquired. A detailed operational description is given in the text.

cascade: If the FM coil current exceeds the threshold level, the clock is applied to the counters in such a way as to change the coarse tuning voltage to result in a reduction of FM coil current magnitude. A 0.5 volt hysteresis for the comparator thresholds provides an overcorrection safety margin.

The FM coil driver and loop filter functions are combined in a single circuit, shown in Figure 8. An LH0022 FET input op-amp is used in conjunction with a diode biased, complimentary-darlington connected, transistor power amplifier to implement the FM coil driver, voltage to current converter circuit. This configuration is capable of delivering currents of ± 0.8 A to bias the FM coil, with current limiting provided by 0.5 ohm emitter resistors at ± 1.0 A. Coil current is sensed by a precision 10 ohm resistor in series with the ground side of the coil. When connected as an amplifier with a gain of 10 V/A (feedback capacitor and damping resistor replaced with 10K) the amplifier small signal bandwidth is greater than 1 MHz and crossover distortion is undetectable at the output for a 5 volt peak to peak sinusoid at the input, for frequencies below 50 KHz. When configured as an integrator for the loop filter function, the 0.01 μ F feedback capacitor and 10K input resistor result in a loop natural frequency of about 150 KHz (phase detector sensitivity, $K_D = 0.25$ V/rad, oscillator sensitivity, $K_O = 7.85 \times 10^5$ rad/sec/mA). As can be seen from Figure 8, the feedback capacitor and damping resistor are shunted by an 18M resistor. Limiting the DC gain (to 1800) in this manner aids in lock acquisition by keeping the coil driver amplifier from latching in a current limiting state with the loop in the unlocked condition. The 0.18 sec time constant introduced by the extra gain limiting resistor in the feedback path does not significantly affect the behaviour of the loop. An offset trim control is provided at the op-amp in




Fig. 8. The circuit diagram of the FM coil driver amplifier which incorporates the loop filter for the YIG oscillator phase locked loop.

order to allow compensation for the DC offset associated with the balanced mixer phase detector.

Figure 4(d) shows the spectrum of the IF signal measured at the IF monitor coupler output with the YIG oscillator phase locked at 30.4500 GHz. The spectrum is free of spurious non-harmonically related components to at least 30 dB below the carrier. The noise sidebands characteristic of a spectrum from a phase locked oscillator with a 3 dB loop bandwidth of about 400 KHz are clearly seen.

III. The 90-120 GHz Frequency Tripler

The frequency tripler employs a split block construction which has been successfully used in a number of different frequency multiplier designs [1],[11]. Details of the multiplier design are given in Figure 9. The tripler is a geometrically scaled version of a design used at 200-290 GHz [11].

Power incident in the full height WR-28 input waveguide is fed to the varactor diode via a tuneable probe transition and a seven-section suspended substrate lowpass filter, which passes the pump frequency with low loss, but is cut-off ($f_c = 56.5$ GHz) for higher harmonics. The transmission line and channel dimensions for the stripline filter were directly scaled by a factor of 1.89 from the higher frequency design [11]. The varactor chip, a 0.1 mm sided cube, is mounted on the filter substrate adjacent to the reduced height output waveguide. One of the many diodes on the chip is contacted and coupled to the output waveguide with a post mounted, 0.31 mm long x 0.0125 mm diameter, phosphor bronze whisker, which has been suitably pointed and prebent. Output tuning is accomplished with the aid of an adjustable backshort in this guide. DC bias is brought to the device by a transmission line filter. The design of the filter has been described extensively in other papers [11] - it will suffice to state that the





Fig. 9. a) A view of the 95-120 GHz tripler block, split along the partition between blocks, showing the input guide, the stripline lowpass filter, the bias circuit and the output waveguide. Dimensions in millimeters.



TRANSFORMER DIMENSIONS

STEP	WIDTH	HEIGHT
0	2.48	0.48
1	2.28	0.61
2	1.84	0.78
3	1.65	0.83

(b)

Fig. 9. b) A section through the block detailing the waveguide transformer and diode mounting arrangement.

bias line approximates a quarter-wave, short circuited stub at 50 GHz, thus minimizing the effect of the bias circuit on the performance of the lowpass filter near cut-off.

A quarter-wave, two section impedance transformer couples the 2.48 mm x 0.48 mm reduced height guide to the 1.66 mm x 0.83 mm output guide. Power can flow in the wider guide at the second harmonic, whereas the output guide is cut off at this frequency. The transformer is thus used to implement a reactive second harmonic idler termination by spacing it approximately $\lambda_g/2$ (at the second harmonic wavelength) from the plane of the diode.

The varactor diode is a Schottky-barrier device fabricated by R. Mattauch at the University of Virginia (designated 8P2) with a zero bias capacitance of 28 fF, a DC series resistance of 10.3 ohms and a breakdown voltage of 20 volts at 1 uA. These devices have a highly non-linear capacitance versus voltage law which approximates the inverse half power behaviour of an ideal abrupt junction varactor to within 5 volts of the breakdown limit.

The performance of the multiplier has been measured at 1500 MHz intervals for output frequencies between 90 and 120 GHz. The output power response as a function of output frequency for constant 50 mW pump power is shown in Figure 10. Backshort tuning and DC bias were optimized at each measurement frequency. Typically the bias conditions for optimum performance were a reverse bias voltage of about 4 volts and a forward current between 1 mA and 4 mA. More than 2 mW output power is obtained at any frequency between 90 and 120 GHz, corresponding to a minimum conversion efficiency, for 50 mW pump power of 4%. The peak output power obtained was 4 mW at 110 GHz, corresponding to an efficiency of 8%.

Figure 11 shows how the multiplier conversion efficiency varies with pump power at three different output frequencies. Bias and tuning were optimized




Fig. 10. Output power versus output frequency for the tripler:
A) at constant 50 mW input power (from a klystron source) and
B) when pumped by the 26.5-40 GHz YIG oscillator.
For curve A), the tripler bias and tuning were optimized at each frequency, whereas
for curve B) the tuning was optimized but the bias was fixed, as described in the text.




Fig. 11. Conversion efficiency versus pump power at three different output frequencies. Bias and tuning were optimized at each measurement power.

for best performance at each pump level. Although the tripler can safely handle input powers of up to 120 mW, the best conversion efficiency occurs at input powers of about 30 mW.

IV. The Complete Synthesized LO Source

Figure 10 shows the output power response of the complete synthesized LO system, with the tripler pumped by the phase locked, YIG-tuned Gunn oscillator. More than 1 mW output power is available at any frequency between 90 and 120 GHz when the tripler tuning is optimized at each operating frequency. The tripler bias was fixed at a voltage of 3.5 volts for these measurements. The tuning curves (position vs. operating frequency) for the tripler backshorts are monotonic and very repeatable. Motorized servo tuning of the multiplier backshorts could easily be implemented automatically by using the frequency select input, resulting in completely automatic operation of the synthesized local oscillator system.

V. Summary

The synthesized 90-120 GHz local oscillator system described in this paper was designed for use with cooled Schottky diode or SIS mixer heterodyne radiometers. The system is fully automatic in operation and exhibits adequate output power, wide tuning range and excellent frequency stability, making it an ideal choice for application as an LO source in millimeter-wave spectrometer receivers. There clearly exist other areas of application for such a signal source. Some possibilities are: i) as a phase stable local oscillator for the receivers in a phase coherent millimeter wave antenna array, ii) as a synthesized swept frequency signal generator for laboratory test applications. The unit described here is to be incorporated in the near future into a 100-120 GHz SIS receiver

system [12] to be installed on the NRAO 12 meter diameter radio telescope near Tucson, Arizona.

VI. Acknowledgments

The authors gratefully acknowledge the assistance of N. Horner Jr. who assembled the tripler block and G. Taylor who fabricated the tripler mount. Prof. R. Mattauch of the University of Virginia is thanked for providing the Schottky diodes used in the multiplier. Thanks also go to M. Balister who provided helpful comments and advice during the construction of the synthesizer.

References

1. J. W. Archer, "All Solid-State Receiver for 210-240 GHz," IEEE Trans on MTT, vol. MTT-30, p. 1247, August 1982.
2. N. R. Erickson, "A High Efficiency Frequency Tripler for 230 GHz," Proc. 12th Eur. Micr. Conf., (Helsinki, Finland), p. 288, Sept 1982.
3. J. E. Carlstrom, R. L. Plambeck and D. D. Thornton, "A 60-115 GHz Tuneable Gunn Oscillator", submitted for publication to IEEE Trans. on MTT, August 1984.
4. C. R. Predmore, N. R. Erickson, P. F. Goldsmith and J. L. R. Marrero, "A Broadband Ultra Low-Noise Schottky Diode Mixer Receiver from 80 to 115 GHz," IEEE Trans. on MTT, vol. MTT-32, p. 498, May 1984.
5. J. W. Archer and M. T. Faber, "A Very Low-Noise Receiver for 80 to 120 GHz," Int. J. I.R. and MM-Waves, vol. 5, no 8, p. 1069, August 1984.
6. S-K. Pan, M. J. Feldman, A. R. Kerr and P. Timbie, "A Low Noise 115 GHz Receiver Using Superconducting Tunnel Junctions," App. Phys. Lett., vol. 43, p. 786, 1983.
7. L. R. D'Addario, "A 90-120 GHz SIS Mixer Having Gain and Wide Bandwidth," Int. J. I.R. and MM-Waves, vol. 5, no. 10, October 1984.
8. F. M. Gardner, "Phaselock Techniques", p. 208, John Wiley, New York, 1979.
9. "Pulse Swallowing Revisited," Fairchild ECL Data Book, p. 9-56/581, Fairchild Camera and Instrument Corp., Mountain View, CA, 1977.
10. F. M. Gardner, ibid, p. 11.
11. J. W. Archer, "An Efficient 200-290 GHz Frequency Tripler Incorporating a Novel Stripline Structure," IEEE Trans. on MTT, vol. MTT-32, p. 416, April 1984.

12. J.W. Archer, "A High Performance, 2.5K Cryostat Incorporating a 100-120 GHz Dual Polarization Receiver", submitted for publication to Rev. Sci. Instr., October 1984.